

Nazwa dokumentu	Symbol	Wydanie	Data wydania
Wzór deklaracji zgodności	A10-TD1	3	2017-08-17

Declaration of Conformity

Manufacturer:

Prodromus Sp. z o.o. (Ltd.)
Golkowice 451
30-698 Krakow, Poland

We declare on their sole responsibility that the medical device: Prodrobot automated gait trainer

has been classified as a class IIa, rule 9 in accordance with the Regulation of the Minister of Health of 5.11.2010r. on the method of classifying medical devices (Journal of laws from 2010, No. 215, item. 1416)

The product meets the requirements of the Directive 93/42 / EEC and the Act of 20 May 2010 on Medical Devices.

The conformity assessment procedure conducted in accordance with Annex II of Directive 93/42/EEC.

All applicable harmonized standards in accordance with the requirements of Directive 93/42 / EEC - can be found in the Technical Documentation.

Notified body



TÜV Rheinland LGA Products GmbH
Tillystraße 2
90431 Nürnberg, Germany

Place and date:

Golkowice, 2017-08-17

Name/Position:

Bartłomiej Wielogórski
C.E.O.

A handwritten signature in blue ink, appearing to read 'Wielogórski', is written over a white rectangular background.